

For questions regarding endoscope channel adaptation or if acquiring new endoscopes, please contact Wassenburg Medical, Inc.

* P/N 16030 is an in-line disc filter that must be replaced as required or every six (6) weeks.

